

The Effect of Free-Field Presentation and Processing Strategy on a Measure of Spectro-Temporal Processing by Cochlear-Implant Listeners

Trends in Hearing
Volume 24: 1–12
© The Author(s) 2020
DOI: 10.1177/2331216520964281
journals.sagepub.com/home/tia



Alan W. Archer-Boyd , Tobias Goehring, and Robert P. Carlyon

Abstract

The STRIPES (Spectro-Temporal Ripple for Investigating Processor Effectiveness) test is a psychophysical test of spectro-temporal resolution developed for cochlear-implant (CI) listeners. Previously, the test has been strictly controlled to minimize the introduction of extraneous, nonspectro-temporal cues. Here, the effect of relaxing many of those controls was investigated to ascertain the generalizability of the STRIPES test. Preemphasis compensation was removed from the STRIPES stimuli, the test was presented over a loudspeaker at a level similar to conversational speech and above the automatic gain control threshold of the CI processor, and listeners were tested using the everyday setting of their clinical devices. There was no significant difference in STRIPES thresholds measured across conditions for the 10 CI listeners tested. One listener obtained higher (better) thresholds when listening with their clinical processor. An analysis of longitudinal results showed excellent test–retest reliability of STRIPES over multiple listening sessions with similar conditions. Overall, the results show that the STRIPES test is robust to extraneous cues, and that thresholds are reliable over time. It is sufficiently robust for use with different processing strategies, free-field presentation, and in nonresearch settings.

Keywords

cochlear implant, spectro-temporal, free-field, behavioral test

Received 19 February 2020; Revised 15 September 2020; accepted 16 September 2020

Most cochlear-implant (CI) listeners understand speech in quiet, though performance can vary greatly. In noisy situations, even high-performing listeners will struggle to understand speech. Many new ways of improving CIs have been developed to increase the number of listeners who might benefit, particularly when listening to speech in noise. The methods range from new stimulation modes (Arora et al., 2011; Bierer & Litvak, 2016; Donaldson et al., 2005; Holden et al., 2002; Litvak, Spahr & Emadi, 2007; Litvak, Spahr, Saoji, et al., 2007; van den Honert & Kelsall, 2007), to novel processing strategies (Goehring et al., 2017, Goehring, Archer-Boyd, et al., 2019; Goehring, Keshavarzi, et al., 2019; Loizou et al., 2000; Riss et al., 2008; van Hoesel et al., 2008) and improvements in the way devices are tailored to individual patients (Garadat et al., 2012; Noble et al., 2014). Suitable tests to quickly and accurately assess these methods have also been sought, as speech tests

rely heavily on a listener's acclimatization to their processing strategy over timescales that are much longer than a clinical appointment (Davis & Johnsrude, 2007; Davis et al., 2005). The acclimatization required with speech tests complicates the evaluation of novel strategies when compared with the listener's clinical strategy, as the latter has a systematic advantage due to long-term usage. The potentially limiting factor of acclimatization has been discussed in several recent studies that used speech to evaluate novel processing strategies (Berg

Cambridge Hearing Group, MRC Cognition and Brain Sciences Unit, University of Cambridge, Cambridge, United Kingdom

Corresponding author:

Alan W. Archer-Boyd, Cambridge Hearing Group, MRC Cognition and Brain Sciences Unit, University of Cambridge, Cambridge CB2 7EF, United Kingdom.

Email: Alan.Archer-Boyd@mrc-cbu.cam.ac.uk



et al., 2019; Croghan et al., 2017; Schwartz-Leyzac et al., 2017). In the worst case, this can lead to false conclusions and the prevention of further investigation into potentially beneficial new methods or strategies. Take-home experience with a new program is a valid solution, but one that, if the program is ultimately unsuccessful, may involve subjecting the listener to long periods of poor hearing.

Many nonspeech tests have been developed in order to circumvent the need for take-home testing, with the aim of providing quick, reliable information on whether a listener's auditory perception has been improved by a manipulation of a CI program (Aronoff & Landsberger, 2013; Azadpour & McKay, 2012; Drennan et al., 2008; Henry & Turner, 2003; Litvak, Spahr, Saoji, et al., 2007; Saoji et al., 2009; Supin et al., 1994; Won et al., 2007, 2011). We have previously described a number of criteria that such tests should meet (Archer-Boyd et al., 2018). These include the necessity for the listener to perform both spectral *and* temporal processing to obtain good performance, and the related requirement that performance should not be possible based on any local spectro-temporal segment. Instead, the test should require a higher order cue to be extracted from the stimuli such as the direction of frequency sweeps (Archer-Boyd et al., 2018). A nonspeech, truly spectro-temporal test avoids the learning effects and translation requirements inherent in speech tests, while measuring an important factor in speech intelligibility, namely, spectro-temporal processing. Here, spectro-temporal processing is defined as requiring a combination of information across both frequency and time, rather than simply requiring the listener to monitor one frequency channel or to perform an instantaneous across-frequency judgment. One test that fulfills these criteria is the spectro-temporal ripple for investigating processor effectiveness (STRIPES) test (Archer-Boyd et al., 2018). The STRIPES test presents listeners with concurrent sinusoids that either decrease or increase in frequency and requires them to discriminate between upward- and downward-sweeping stimuli (cf. Figure 1). Difficulty is

increased by increasing the temporal overlap between each sweep. Although the amplitude modulation (AM) in any one frequency region changes with stripe density, it remains the same for both up and down sweeps. The STRIPES test has been shown to be sensitive to large reductions in spectral resolution at the level of individual listeners, and the test was simple enough to obtain thresholds from newly implanted listeners (Archer-Boyd et al., 2018). The test also successfully predicted which of two experimental fitting algorithms, each of which deactivated different subsets of electrodes, produced the best speech-in-noise performance at the level of individual listeners (Goehring Archer-Boyd, et al., 2019; Goehring, Keshavarzi, et al., 2019) and was correlated with an individual measure of channel interaction based on spectrally blurred speech-in-noise stimuli (Goehring et al., 2020).

STRIPES was initially developed for the HiRes-S strategy implemented in CIs manufactured by Advanced Bionics, and great care was taken to remove any within-channel cues (at least, those visible in the electrodiagrams) that could be used to perform the task based on cues other than spectro-temporal modulation. Stimuli were presented at a level below the automatic gain control threshold (AGC) via a direct-audio input from a soundcard to the CI processor, removing possible confounding factors such as room reverberation, loudspeaker/microphone frequency response, preemphasis processing, and spectral filtering due to head movement. As much additional processing of the stimuli (e.g., noise reduction) was disabled as possible. In comparison, the majority of research and clinical auditory tests are presented in free-field via a loudspeaker, at a sound level similar to normal conversational speech (60–70 dB SPL, above the threshold of most AGCs), and identical stimuli are used across manufacturers (e.g., Aronoff & Landsberger, 2013; Kollmeier et al., 2015; Spahr et al., 2012). Input is via a microphone on the CI processor, and the listener's everyday clinical program is often used, meaning that additional processing such as




Figure 1. Spectrograms of the STRIPES Stimuli at Density = 5 (1 RPO/5 Hz AM). The left plot shows up STRIPES, and the middle and right panels show two down STRIPES at different starting phases.

noise-reduction algorithms may remain active and affect the presented stimuli.

All of these factors make testing easier in the clinical setting, but potentially introduce extraneous cues that could be used to perform the task. Indeed, spectral and spectro-temporal ripple tests such as the spectral-temporally modulated ripple test by Aronoff and Landsberger (2013) can contain cues that allow the task to be performed using either purely within-channel temporal processing or by comparing the across-channel profile of excitation in a single short time segment—hence not meeting our proposed requirement for truly spectro-temporal processing. As noted earlier, a spectro-temporal cue is defined as any cue that requires comparison across time *and* frequency. Nonspectro-temporal cues, which require only temporal *or* spectral processing, could include within-channel AM cues at low ripples-per-octave (Anderson et al., 2012; Archer-Boyd et al., 2018), and spectral aliasing due to the number and spacing of the filter bands used in the vocoder or CI (conference abstract, O'Brien & Winn, 2017). These issues are less of a problem for the STRIPES test, for which psychometric functions are monotonic for CI listeners (Archer-Boyd et al., 2018). A study using stimuli and methods that are similar to the STRIPES test has found no evidence of nonspectro-temporal cues and has found that performance can be predicted by the highest performing octave band, not by average broadband performance in normal-hearing and hearing-impaired listeners (Narne et al., 2018).

The work presented here quantifies the effect on listener performance of easing the previously strict experimental conditions and thus potentially making the test more flexible and applicable to a clinical setting. To do so, the effect of presenting the stimuli via loudspeaker instead of the direct auditory input was examined. The use of the patient's own clinical processing strategy—rather than the HiRes-S strategy used for developing STRIPES—was also examined.

All listeners used the HiRes Optima-S clinical strategy in their everyday settings. In previous STRIPES studies, the experimental maps used the HiRes-S strategy. One major difference between these strategies is that Optima uses current steering between electrodes. Current steering potentially increases the number of channels (or virtual channels) that a listener can discriminate (Firszt et al., 2007; Koch et al., 2007; Landsberger & Srinivasan, 2009). However, speech intelligibility results have been mixed when using current steering across the whole array versus no steering (Firszt et al., 2009; Mens & Berenstein, 2005). The Optima-S strategy limits current steering such that the proportion of current applied to the more apical electrode of each pair ranged between 0.25 and 0.75 (Advanced Bionics, 2013).

This study compares the performance on different implementations of the STRIPES test. One of these is the original method using direct auditory input and the HiRes-S strategy, reported by Goehring, Archer-Boyd, et al. (2019), whose data are reproduced here. The second uses the same HiRes-S strategy but presents the stimuli via loudspeaker. Finally, loudspeaker presentation is used, combined with the listener's clinical (HiRes Optima-S) strategy. The data from all three conditions were obtained from the same listeners.

Methods

Overview of the STRIPES Test

The STRIPES test procedure is identical to the implementation described by Goehring, Archer-Boyd, et al. (2019) and is described briefly here. The test uses an adaptive procedure to measure the spectral-density threshold at which the listener can just distinguish the target stimulus from two reference stimuli in a three-interval, two-alternative forced-choice task. Stimuli are created from 1 s-long, concurrent, and concatenated exponential sine sweeps (with random starting phase) moving up or down in frequency from 250 to 8000 Hz, at a rate of 5 octaves per second. The delay between the start of each sweep is $1/\text{density}$, such that at density = 5, the delay between the start of each sweep is 200 ms. As the bandwidth of the stimulus is 5 octaves, the number of ripples per octave (RPO) is equal to density/5. As the individual sweeps are 1 s in duration, the AM frequency in Hz is the same as the density. A single cycle of a STRIPES stimulus can be described in the discrete time domain by the following equation, after Berdahl and Smith (2008),

$$s(n) = \sum_{d \in \mathbb{Z} \cap [0, D]} \sin \left[K \left(e^{-\frac{(n + \frac{df_s}{D})}{L f_s}} - 1 \right) \right]$$

$$\text{if } n + \left(\frac{df_s}{D} \right) \leq T f_s, \text{ for } n = 1, \dots, \frac{f_s}{D}$$

where n = sample, f_s = sample rate (=48kHz), D = density, d = individual sweep, T = sweep duration (=1 s), $K = \frac{\omega_1 T}{\ln(\frac{\omega_2}{\omega_1})}$, and $L = \frac{T}{\ln(\frac{\omega_2}{\omega_1})}$.

The number of cycles, and hence the duration of the stimuli, decreases with increasing density. This was in order to ensure that at least two unbroken single sweeps are presented in each stimulus regardless of the starting phase, and that an integer number of cycles at each density is presented. The duration of the STRIPES stimuli varied from 2.07 s (density = 1.1) to approximately 1.3 s at the highest densities presented. Ensuring an integer number of cycles also has the effect of matching

the start and end instantaneous frequencies for each stimulus, minimizing this potentially confusing and salient cue. The stimuli used for loudspeaker presentation were identical to those used in the normal hearing (NH) vocoder experiment (Experiment 1) of Archer-Boyd et al. (2018). The changes from the previous CI version of the stimuli were the removal of preemphasis compensation, and the application of equal 50-ms onset/offset ramps to each individual sweep in the stimuli.

The preemphasis compensation was initially added to the stimuli to reduce the effect of the high-pass preemphasis filter used in AB devices and described in Boyle et al. (2009). However, preemphasis does not introduce any additional cues that would allow the listener to perform the STRIPES task and may result in the low-frequency channels being higher in level than in everyday listening. Therefore, in this version, preemphasis compensation was removed. Previously, the duration of the ramps imposed to smooth the onsets and offsets of the glides differed depending on whether each glide ended at a high (8000 Hz) or a low (250 Hz) frequency. This was done in order to make the individual sweep envelopes more symmetric in the lowest and highest channels where the sweeps started and ended. However, this symmetry changed depending on the filters used and would not be practical for testing across strategies, so standard, equal-duration (50 ms) raised cosine ramps were applied to the starts and ends of all sweeps.

The listener selects the target interval, which is either the first or last interval, and which is always an upward sweep; the other two intervals contain downward sweeps. The starting phase of the stimuli is randomized. The number of concurrent frequency sweeps (the density) is varied to titrate difficulty, with the task being easy at a density close to one peak per 5 octave sweep, and progressively harder at higher densities. Note that non-integer density values are possible; for example, a density of 2.5 would mean that 50% of the time two swept sinusoids were present simultaneously (overlapped) and that for the other 50% of the time three swept sinusoids overlapped. Example spectrograms of the stimuli at density 5 (1 RPO and 5 Hz AM) are shown in Figure 1. The starting frequency is roved across trials and the beginning

and end of each interval are masked by short noise bursts to reduce the salience of onset and offset cues. An adaptive two-up/one-down procedure starts with a sweep density of 1.1 (number of sweeps concurrently presented during each sweep) and adjusts the density per trial with a density step size of 0.5 (for the first four reversals) and 0.2 (for the last eight reversals). The interstimulus interval was 0.6 s. The test ends after 12 reversals and the final threshold of the run was calculated as the average of the last 4 reversals.

Apparatus

Listeners took part in the STRIPES test in a double-walled Industrial Acoustics Company (IAC Acoustics, Winchester, UK) sound-proof booth measuring $2 \times 2 \times 2$ m. A Dell Inspiron laptop (Dell Inc., Round Rock, TX) was used to present the stimuli through an RME Fireface UCX (RME, Haimhausen, Germany) external soundcard connected to a Genelec 8030 C loudspeaker (Genelec Oy, Iisalmi, Finland). The loudspeaker had a flat (± 2 dB) response from 250 to 8000 Hz (the bandwidth of the stimuli). The loudspeaker was placed 0.75 m from the wall opposite the door, and 0.35 m from the wall adjacent to the door, at a height of 1.2 m to the center of the loudspeaker. The loudspeaker was deliberately placed off-axis in the room, in order to minimize the effect of room modes. An office chair was placed in front of the speaker, such that the listener's ear was approximately 1 m from the loudspeaker. The stimuli were presented at a level of 65 dB SPL. This level was calibrated using a sound-level meter 1 m from the loudspeaker at a height of 1.2 m, using a noise with the same long-term spectrum and root mean square (RMS) as the STRIPES stimuli. As a final check, the levels of up and down STRIPES at a density of 10 were also measured at the same position.

Conditions

Table 1 shows the differences between all experimental conditions. Clin. HiRes Aux 1 & 2 were identical conditions, tested in two different sessions. Clinical (Clin) means the processor is the same as those used in the

Table 1. Differences Between the Experimental Conditions.

| Condition | Device/CI program | Presentation | Preemphasis compensation | Sweep | | No. participants |
|--------------------------|---|----------------------|--------------------------|-------------------------|--|------------------|
| | | | | onset/offset ramps (ms) | | |
| Clinical HiRes Aux 1 & 2 | Clinical Harmony TM /HiRes-S | Auxiliary input | Yes | 100/25 | | 8 |
| Research HiRes Aux | Research Harmony TM /HiRes-S | Auxiliary input | Yes | 100/25 | | 10 |
| Research HiRes LS | Research Harmony TM /HiRes-S | Loudspeaker/omni-mic | No | 50/50 | | 10 |
| Clinical Optima LS | Clinical Naida TM /Optima-S | Loudspeaker/T-mic | No | 50/50 | | 10 |

clinic and programmed using the clinical fitting program Soundwave 3 (Advanced Bionics, 2019). High resolution (HiRes) is AB's older continuously interleaved sampling (CIS) strategy, without current steering. Auxiliary (Aux) input is the direct audio input on the Harmony™ that bypasses the microphone input and presents audio directly from a soundcard. Rsrch. HiRes Aux was the same condition tested using a research Harmony™ processor instead of a clinical Harmony™ processor (see Table 3 for the differences between the devices). The research processor was programmed using the Bionic Ear Programming System Plus, or Bionic Ear Programming System plus (BEPS+) (Advanced Bionics, 2014). The main difference between the research and clinical Harmony™ is the filters, and these will be described in greater detail in the hardware and processing differences section. The results from these conditions were previously presented in Goehring, Archer-Boyd, et al. (2019). Rsrch. HiRes LS used a research Harmony™ processor and a different version of the STRIPES stimuli (detailed in Overview of the STRIPES test, Methods), presented over a loudspeaker (LS). Clin. Optima LS used the same presentation method and test stimuli as Res. HiRes LS, and the listener used their clinical device. Optima is the current-steering CIS strategy currently used by default in clinical AB devices. The strategy uses pairs of electrodes for each channel, delivering between 25% and 75% of the current to each electrode in a pair.

Listener Demographics and CI Processor Settings

Ten experienced CI listeners, all of whom were unilateral users of an Advanced Bionics implant, took part. The demographic information of the listeners is given in Table 2. Ethical approval was obtained from the National Research Ethics Committee for the East of England. Before commencing the experiments listeners gave their informed consent to participate and were informed that they could withdraw from the study at any point. Listeners were paid for taking part and travel expenses were reimbursed. Listeners' clinical settings varied, with different levels of ClearVoice™ noise reduction applied. ClearVoice™ aims to identify channels that do not contain speech energy and reduce the gain applied to them in order to enhance the signal-to-noise ratio (Advanced Bionics, 2012). One listener used a linked contralateral routing of signal device. All listeners' everyday clinical programs were HiRes Optima-S™.

Test Procedure

Listeners were seated in the booth, and pseudo-randomly assigned to first listen through either their own clinical

processor set to everyday program (Clin. Optima LS) or through the Research Harmony™ processor, running a HiRes version of their map with no further signal processing enabled (Rsrch. HiRes LS). Half started with their clinical device. The levels of the STRIPES stimuli were gradually increased to the required 65 dB SPL. The listeners were asked to rate the level of the stimuli on an 11-point loudness scale ranging from 0 (inaudible) to 10 (uncomfortably loud). A rating of 6 (most comfortable) was required. To obtain this rating for listeners AB3 and AB19, the level had to be further increased by adjusting the volume control on the device. The STRIPES test was then explained to the listeners, and the experimenter exited the booth and closed the doors before the test was started by the listener.

The listeners were presented with five pretest forced-choice trials, at density = 1.1 (easiest), in order to familiarize themselves with the STRIPES test. The pretest was repeated until the listener achieved at least 4/5 correct trials; all listeners achieved this on their first attempt. After successful completion of the pretest, the listeners completed one run of the adaptive track and raised their hand to indicate that the test was complete. The experimenter reentered the booth and noted the threshold for the run. After two runs, if the thresholds were different by >1 density, a third run was performed, and the scores from all three were averaged. After completion of two or three runs with one device and program setting, the listener was fitted with the other device and program setting, and the same procedure was carried out. In total, between four and six thresholds were measured for each listener in their first and/or only session. Each run took approximately 8.5 min (*SD* 1.3 min). The duration of each session was 45 min to 1 hr.

Hardware and Processing Differences

The experimental conditions differed both in the processing strategy used and in the processor hardware, which was either a laboratory-owned clinical or research Harmony™ processor in the HiRes-S conditions, or the listener's own Naida™ processor in the Optima-S conditions. Key differences and similarities between the three device/processing types are shown in Table 3. One specific difference to note concerns the filters used in the clinical Harmony™ versus the research Harmony™ and clinical Naida™ processors. The frequency domain filters in the research Harmony™ and clinical Naida™ processors use a 128-point Fourier transform, which is separated into bands for envelope extraction. Electrical output of STRIPES when viewed on an oscilloscope was markedly different between the two devices in the low frequencies. The time-domain filters were sharper than the frequency domain filters

Table 2. Listener Demographic Information.

| Listener | Sex | Age (years) | Duration implanted | Duration since onset of profound hearing loss | Etiology/ acquired pre-/postlingual | CI speech processor | Implant type/CI electrode array | Clinical strategy | Pulse width (μ s) | Electrodes deactivated |
|----------|-----|-------------|--------------------|---|---|---------------------|---------------------------------|--|------------------------|------------------------|
| | | | | | | | | | | |
| AB1 | M | 74 | 10 | 41 | Unknown/postlingual | Naida CI Q90 | HR90K/HiFocus I J | HiRes Optima-S/ ClearVoice medium | 26.0 | 16 |
| AB2 | F | 60 | 11 | 27 | Possible ototoxicity/ postlingual | Naida CI Q90 | HR90K/ HiFocus I J | HiRes Optima-S/ ClearVoice medium | 35.0 | 16 |
| AB3 | M | 72 | 11 | 36 | Otosclerosis/postlin- gual progression | Naida CI Q90 | HR90K/ HiFocus I J | HiRes Optima-S/ ClearVoice off/CRoS | 30.5 | 12 |
| AB6 | F | 70 | 5 | 65 | Unknown/peri-lingual | Naida CI Q90 | HR90K/ HiFocus I J | HiRes Optima-S/ ClearVoice high | 35.0 | 16 |
| AB19 | M | 75 | 3 | - | Unknown/postlingual progression | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S/ ClearVoice medium/ hearing aid removed | 34.1 | None |
| AB20 | M | 74 | 3 | >30 | Unknown/postlingual progression | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S/ ClearVoice medium | 29.6 | None |
| AB23 | F | 60 | 3 | 58 | Unknown/peri-lingual | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S/ ClearVoice medium | 24.2 | None |
| AB24 | F | 49 | 3 | 4 | Unknown/postlingual sudden | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S/ ClearVoice off | 34.1 | 15, 16 |
| AB25 | F | 66 | 3 | - | Unknown/postlingual | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S/ ClearVoice medium | 18 | 16 |
| AB26 | F | 58 | 5 | - | Unknown/postlingual | Naida CI Q90 | HR90k Advantage/ HiFocus MS | HiRes Optima-S | 22.4 | None |

Note. CI = cochlear-implant; CROS = contralateral routing of signal.

Table 3. Comparison of Clinical and Research Harmony™ With HiRes-S and Naida™ With Optima-S.

| Hardware/ Processing | Clinical Harmony™/HiRes-S | Research Harmony™/HiRes-S | Clinical Naida™/Optima-S |
|--------------------------------------|---|---|---|
| Input/Microphone | Auxiliary input; Omni-dir.; Front behind-the-ear (BTE)-type | Auxiliary input; Omni-dir.; Front behind-the-ear (BTE)-type | Omni-dir.; in-ear T-mic |
| Channel filters | Time domain | Frequency domain; Fourier transform (FT) based | Frequency domain; Fourier transform (FT) based |
| Current steering Max. Voltage (V) | None 8 | None 8 | Max. 75%/Min. 25% 4 |




Figure 2. Mean and Single STRIPES Thresholds for Individual Participants. The black bars and squares show data from the original STRIPES test with a HiRes-based program and the same device across participants, reproduced from Goehring, Archer-Boyd, et al. (2019). Red bars/squares show the mean/single thresholds for the same device/program as before, but with STRIPES presented over a loud-speaker. Blue bars/squares show thresholds measured using the same loudspeaker-presented STRIPES test and the participant’s clinical device, set to their everyday program. Black and blue triangles show thresholds collected during additional sessions investigating AB24’s high performance and are not included in the average and standard errors.

in the first two to three channels, depending on frequency allocation. This parameter may have affected STRIPES performance as the filter roll-off changed the shape (essentially, the AM) of the envelope in the low-frequency channels. The Naida™ uses a sequential, current-steered CIS strategy (Optima-S™), whereas the Harmony does not use current steering (HiRes-S™). Current steering means that all electrodes stimulate in pairs. In the Optima strategy, it is constrained so that at least 25% of the total charge is presented to one electrode, and no more than 75% of the charge presented to the other.

Results

Data from the free-field HiRes-S and Optima S conditions are shown by the red and blue bars, respectively, in Figure 2; data from Goehring, Archer-Boyd, et al. (2019) obtained with the previous version of the STRIPES test (e.g., preemphasis compensation, presentation below AGC threshold, etc.), using the direct (i.e., auxiliary) input and HiRes-S strategy are shown by the black bars for comparison. A one-way repeated-measures analysis of variance showed no effect of condition, $F(2, 27) = 0.31, p = .7343$. Hence

there is no indication that, overall, changing to loudspeaker presentation or using the patient's clinical device and strategy either improved or degraded performance. One listener, AB24, performed markedly better with their clinical Naida processor and Optima-S strategy. AB26 showed a smaller possible improvement. Subsequent sessions to find the source of AB24's high performance were inconclusive (not shown).

Discussion

Comparability and Replicability of STRIPES Implementations

As noted in the Results section, there was no significant difference in overall performance between the three different implementations of the STRIPES test. Of greater interest is the extent to which important information about the variation in spectro-temporal sensitivity is preserved when switching to loudspeaker presentation and from the HiRes-S to the clinical strategy. One

common method for assessing agreement between two clinical tests is to measure the across-listener correlation between them. A high correlation shows that the two tests capture the same underlying variability across listeners but does not necessarily show that they agree—for example, one test could consistently provide thresholds double that of the other. An alternative method was recommended by Bland and Altman (1986). That method is first illustrated here with reference to the data of Goehring, Archer-Boyd, et al. (2019) who measured STRIPES thresholds for three maps twice on two different days—essentially a test-retest experiment. This provides a “gold standard” against which to test agreement between conditions ($N=8$). Depicted in Figure 3A, the across-listener Pearson correlation between the two sessions (Clin. HiRes Aux 1 and 2) was strong ($r=.96$, $p<.001$, $df=7$). Bland and Altman recommend first plotting the difference between the two sessions as a function of their mean, which is done in Figure 3D, showing that the error is roughly constant across different levels of performance ($r=.03$, $p=.94$, $df=7$) and that




Figure 3. STRIPES Performance for Three Comparisons. The top row shows the correlations between thresholds in each pair of conditions with regression lines (solid lines), and the bottom row shows the differences between thresholds plotted against the average in thresholds with the mean of differences (solid lines). (A and D) Test-retest results based on data from Goehring, Archer-Boyd, et al. (2019) who used the same experimental setup and test conditions and serves as baseline ($N=8$). (B and E) Comparison of STRIPES thresholds between auxiliary and loudspeaker presentation using a research HarmonyTM processor ($N=10$). (C and F) Comparison of STRIPES thresholds between a research HarmonyTM processor with a HiRes program and a clinical NaidaTM processor with an Optima program, using loudspeaker presentation. The dotted lines on the bottom row plots indicate the \pm two SDs from the mean of differences.

there is no substantial bias (mean difference = 0.09). They note that 95% of points should fall within ± 1.96 standard deviations (*SD*), which is indicated by the dotted lines and where the *SD* corresponds to a density of 0.49. It can also be seen that 95% of points show an absolute deviation of less than 1 density unit. This provides some support for the practice of obtaining a third run when the density thresholds from the first two differ by more than density = 1 (0.2 RPO/1 Hz AM).

Performance in the HiRes-S condition with direct input versus loudspeaker presentation is shown in Figure 3B (Rsrch. HiRes Aux and Rsrch. HiRes LS) and correlated strongly for the listeners in this experiment ($r = .91$, $p < .001$, $df = 9$). The Bland–Altman method (Figure 3E) shows no evidence that the error covaried with the mean ($r = -.12$, $p = .74$, $df = 9$) nor that there was a bias (mean difference = 0.24), and a *SD* of 0.72 which is only moderately higher than the test–retest reliability (Figure 3D). The correlation between the two loudspeaker conditions (Rsrch. HiRes LS and Clin. Optima LS) was $r = .91$ when listener AB24's data were excluded and $r = .94$ when they were included (with $p < .001$, $df = 9$, for both). Again, there was no evidence that the error correlated with overall

performance (Spearman's $\rho = -0.6$, $p = .07$, $df = 9$; the error was not normally distributed as determined by the Anderson–Darling test with $p < .001$). There was also no evidence for a substantial bias (mean difference of -0.35 when excluding AB24 and -0.73 for all listeners). The *SD* of errors was 0.95 without listener AB24 and 1.5 for all listeners (Figure 3F).

Overall, the STRIPES test appears robust to manipulations that make it easier to use clinically; a common pattern of variation is observed across listeners, and there is no consistent increase or decrease in performance. Hence, it appears that both the overall level of performance and the variation across listeners are generally impervious to cues that may arise from the AGC, processing strategy, room effects, loudspeaker/microphone frequency response, and spectral filtering due to (small) head movements. However, it is important to understand what cues are available to the listeners in each condition. The next subsection performs such a comparison for the HiRes-S and Optima-S processing strategies. This comparison is of particular interest given that the largest difference observed between any pair of conditions in the experiment was between these two processing strategies, with listener AB24 showing unusually high performance with the Optima S strategy.




Figure 4. Electrodegram Plots of Density = 5 STRIPES Stimuli When Processed Using AB24's HiRes (Left Panel) and Optima (Right Panel) Programs. The top row shows UP STRIPES, and the bottom row shows DOWN STRIPES.

Comparison Between HiRes-S and Optima-S Conditions

The left-hand column of Figure 4 shows electrodiagrams for the HiRes-S strategy at a density of 5, for up glides (top row) and down glides (bottom row). This was produced using the BEPS+ MATLAB simulator, meaning that it uses the research HarmonyTM frequency domain filters. This density is close to the average threshold across all listeners shown in Figure 2 (color online). The directions of the glides are clearly visible and, importantly, there is no clear cue in any one channel that would allow the listener to perform the task. Analogous plots for the Optima-S strategy are shown in the right-hand column. Again, the directions of the glides are visible. A difference between the two strategies is that the envelope peaks are sharper for Optima-S than for HiRes-S. One reason why this did not lead to a difference in performance for most listeners is that the spread of excitation is probably more blurred than the electrodiagrams, due to current spread within the cochlea and, possibly, by neural factors.

It can also be seen that a local cue is introduced by the Optima-S strategy: The envelopes in channels 1–3 and 14 are asymmetric, and the direction of the asymmetry is different for up and down glides. Hence, in principle, the Optima-S strategy introduces a within-channel cue that listeners could have exploited. Unfortunately, attempts to remove the cue by setting the T and C levels in these edge channels to zero were only partially successful, with the asymmetry shifting to the new edge channels. Further attempts with listener AB24 to mask the asymmetries using modulated sine tones produced inconclusive results. During these additional listening sessions, the initial conditions were also retested (Figure 2, black and blue triangles). The repeat measurements of the Optima-S condition revealed much higher variability. There is the possibility that the cues AB24 is using at high thresholds are not related to the monotonically changing density/RPO cue, but to some other nonmonotonic cues that appear due to spectral aliasing of the stimuli due to the bandwidths of the filters used (conference abstract, O'Brien & Winn, 2017). It is also possible that higher nerve density in the cochlea of AB24 and AB26 enabled them to take advantage of the virtual channels provided by OptimaTM current steering. Previous studies have shown that current steering improves the performance in spectral ripple tasks for some listeners (Drennan et al., 2010). However, there is too little evidence in this study to ascertain exactly what cues AB24 was using. In addition, the electrodiagrams do not show the interactions of current or the change in neural response when current steering is enabled, so observations of behavioral results combined with electrodiagrams are necessarily limited.

The fact that performance with the two strategies was similar for most listeners suggests that they did not use the cues available to AB24, although one cannot completely exclude the possibility that Optima-S somehow degraded the true spectro-temporal cue, and that use of the within-channel cue more-or-less exactly compensated for this. The improvement observed for AB24 may have been due to their overall high level of performance: Their excellent spectro-temporal resolution may have allowed them to exploit the improvement in the spectro-temporal representation of Optima-S or to have honed in on the extraneous within-channel cues.

Conclusions

The effect of loudspeaker presentation, within- and across-channel envelope asymmetry cues, and processing strategy on the STRIPES test were investigated. Previous results were compared with results obtained from the same listeners using the new loudspeaker-presented version of the test. Listeners were tested using the same device and map as previously, and using their clinical devices and everyday program. No overall effect of presentation/test type or device/processing type was found. Therefore, the STRIPES test was robust to such manipulations and the nonspectro-temporal cues that were introduced such as the asymmetrical channel envelopes observed in the electrodiagrams (temporal only) and preemphasis (spectral-only weighting toward responding based on higher frequency channels). However, one listener obtained higher and more variable STRIPES thresholds using their clinical device relative to the other conditions tested. The source of this improved performance could not be clearly determined. A comparison of all published STRIPES thresholds to date showed that STRIPES produces generally robust and consistent thresholds for CI listeners across test sessions. The STRIPES test can be used as both a research and clinical tool to investigate listeners' spectro-temporal resolution over a wide range of listening performance.

Acknowledgments

The authors thank François Guérit and John Deeks for their helpful comments on AB24's high performance.

Declaration of Conflicting Interests

The authors declare no potential conflicts of interest with respect to the research, authorship, and/or publication of this article.

Funding

The authors disclosed receipt of the following financial support for the research, authorship, and/or publication of this article:

This work was supported by award G101400 (UK Medical Research Council) to author R.P.C.

References

- Advanced Bionics. (2012). *ClearVoice by advanced bionics clinical results*. Retrieved from: https://advancedbionics.com/content/dam/advancedbionics/Documents/Global/en_ce/Products/Technology/Clearvoice%20White%20Paper.pdf
- Advanced Bionics (2013). Optima clinical results. Retrieved from: https://advancedbionics.com/content/dam/advancedbionics/Documents/Global/en_ce/Professional/Technical-Reports/Sound-Processing/HiRes-Optima-White-Paper.pdf
- Advanced Bionics (2019). Soundwave 3.2, Valencia, CA.
- Anderson, E. S., Oxenham, A. J., Nelson, P. B., & Nelson, D. A. (2012). Assessing the role of spectral and intensity cues in spectral ripple detection and discrimination in cochlear-implant users. *The Journal of the Acoustical Society of America*, *132*, 3925–3934. <https://doi.org/10.1121/1.4763999>
- Archer-Boyd, A. W., Southwell, R. V., Deeks, J. M., Turner, R. E., & Carlyon, R. P. (2018). Development and validation of a spectro-temporal processing test for cochlear-implant users. *The Journal of the Acoustical Society of America*, *144*(5), 2983–2997. <https://doi.org/10.1121/1.5079636>
- Aronoff, J. M., & Landsberger, D. M. (2013). The development of a modified spectral ripple test. *The Journal of the Acoustical Society of America*, *134*(2), EL217–222. <https://doi.org/10.1121/1.4813802>
- Arora, K., Vandali, A., Dowell, R., & Dawson, P. (2011). Effects of stimulation rate on modulation detection and speech recognition by cochlear implant users. *International Journal of Audiology*, *50*(2), 123–132. <https://doi.org/10.3109/14992027.2010.527860>
- Azadpour, M., & McKay, C. M. (2012). A psychophysical method for measuring spatial resolution in cochlear implants. *Journal of Association for Research in Otolaryngology*, *13*(1), 145–157. <https://doi.org/10.1007/s10162-011-0294-z>
- Berdahl, E. J., & Smith III, J. O. (2008). *Transfer function measurement toolbox*. https://ccrma.stanford.edu/realsim/ple/imp_meas/
- Berg, K. A., Noble, J. H., Dawant, B. M., Dwyer, R. T., Labadie, R. F., & Gifford, R. H. (2019). Speech recognition as a function of the number of channels in perimodiolar electrode recipients. *The Journal of the Acoustical Society of America*, *145*(3), 1556–1564. <https://doi.org/10.1121/1.5092350>
- Bierer, J. A., & Litvak, L. (2016). Reducing channel interaction through cochlear implant programming may improve speech perception: Current focusing and channel deactivation. *Trends in Hearing*, *20*, 2331216516653389. <https://doi.org/10.1177/2331216516653389>
- Bland, J. M., & Altman, D. G. (1986). Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet*, *1*(8476), 307–310. [https://doi.org/10.1016/S0140-6736\(86\)90837-8](https://doi.org/10.1016/S0140-6736(86)90837-8)
- Boyle, P. J., Buechner, A., Stone, M. A., Lenarz, T., & Moore, B. C. J. (2009). Comparison of dual-time-constant and fast-acting automatic gain control (AGC) systems in cochlear implants. *International Journal of Audiology*, *48*(4), 211–221. <https://doi.org/10.1080/14992020802581982>
- Croghan, N. B. H., Duran, S. I., & Smith, Z. M. (2017). Re-examining the relationship between number of cochlear implant channels and maximal speech intelligibility. *The Journal of the Acoustical Society of America*, *142*(6), E1537–E1543. <https://doi.org/10.1121/1.5016044>
- Davis, M. H., & Johnsrude, I. S. (2007). Hearing speech sounds: Top-down influences on the interface between audition and speech perception. *Hearing Research*, *229*(1–2), 132–147. <https://doi.org/10.1016/j.heares.2007.01.014>
- Davis, M. H., Johnsrude, I. S., Hervais-Adelman, A., Taylor, K., & McGettigan, C. (2005). Lexical information drives perceptual learning of distorted speech: Evidence from the comprehension of noise-vocoded sentences. *Journal of Experimental Psychology General*, *134*(2), 222–241. <https://doi.org/10.1037/0096-3445.134.2.222>
- Donaldson, G. S., Krefth, H. A., & Litvak, L. (2005). Place-pitch discrimination of single-versus dual-electrode stimuli by cochlear implant users. *The Journal of the Acoustical Society of America*, *118*(2), 623–626. <https://doi.org/10.1121/1.1937362>
- Drennan, W. R., Longnion, J. K., Ruffin, C., & Rubinstein, J. T. (2008). Discrimination of Schroeder-phase harmonic complexes by normal-hearing and cochlear-implant listeners. *Journal of Association for Research Otolaryngology*, *9*(1), 138–149. <https://doi.org/10.1007/s10162-007-0107-6>
- Drennan, W. R., Won, J. H., Nie, K., Jameyson, E., & Rubinstein, J. T. (2010). Sensitivity of psychophysical measures to signal processor modifications in cochlear implant users. *Hearing Research*, *262*(1–2), 1–8. <https://doi.org/10.1016/j.heares.2010.02.003>
- Firszt, J. B., Holden, L. K., Reeder, R. M., & Skinner, M. W. (2009). Speech recognition in cochlear implant recipients: comparison of standard HiRes and HiRes 120 sound processing. *Otology & Neurotology*, *30*(2), 146–152. <https://doi.org/10.1097/MAO.0b013e3181924ff8>
- Firszt, J. B., Koch, D. B., Downing, M., & Litvak, L. (2007). Current steering creates additional pitch percepts in adult cochlear implant recipients. *Otology & Neurotology*, *28*(5), 629–636. <https://doi.org/10.1097/01.mao.0000281803.36574.bc>
- Garadat, S. N., Zwolan, T. A., & Pfungst, B. E. (2012). Across-site patterns of modulation detection: Relation to speech recognition. *The Journal of Acoustical Society of America*, *131*(5), 4030–4041. <https://doi.org/10.1121/1.3701879>
- Goehring, T., Archer-Boyd, A., Deeks, J. M., Arenberg, J. G., & Carlyon, R. P. (2019). A Site-selection strategy based on polarity sensitivity for cochlear implants: Effects on spectro-temporal resolution and speech perception. *Journal of Association for Research in Otolaryngology*, *20*(4), 431–448. <https://doi.org/10.1007/s10162-019-00724-4>
- Goehring, T., Arenberg, J. G., & Carlyon, R. P. (2020). Using spectral blurring to assess effects of channel interaction on speech-in-noise perception with cochlear implants. *Journal for the Association for Research in Otolaryngology*, *21*, 353–371. <https://doi.org/10.1007/s10162-020-00758-z>
- Goehring, T., Bolner, F., Monaghan, J. J., van Dijk, B., Zarowski, A., & Bleeck, S. (2017). Speech enhancement

- based on neural networks improves speech intelligibility in noise for cochlear implant users. *Hearing Research*, 344, 183–194. <https://doi.org/10.1016/j.heares.2016.11.012>
- Goehring, T., Keshavarzi, M., Carlyon, R. P., & Moore, B. C. J. (2019). Using recurrent neural networks to improve the perception of speech in non-stationary noise by people with cochlear implants. *The Journal of the Acoustical Society of America*, 146(1), 705–718. <https://doi.org/10.1121/1.5119226>
- Henry, B. A., & Turner, C. W. (2003). The resolution of complex spectral patterns by cochlear implant and normal-hearing listeners. *The Journal of the Acoustical Society of America*, 113(5), 2861–2873. <https://doi.org/10.1121/1.1561900>
- Holden, L. K., Skinner, M. W., Holden, T. A., & Demorest, M. E. (2002). Effects of stimulation rate with the Nucleus 24 ACE speech coding strategy. *Ear and Hearing*, 23(5), 463–476. <https://doi.org/10.1097/00003446-200210000-00008>
- Koch, D. B., Downing, M., Osberger, M. J., & Litvak, L. (2007). Using current steering to increase spectral resolution in CII and HiRes 90K users. *Ear and Hearing*, 28(2), 38s–41s. <https://doi.org/10.1097/AUD.0b013e31803150de>
- Kollmeier, B., Warzybok, A., Hochmuth, S., Zokoll, M. A., Uslar, V., Brand, T., & Wagener, K. C. (2015). The multi-lingual matrix test: Principles, applications, and comparison across languages: A review. *International Journal of Audiology*, 54 Suppl 2, 3–16. <https://doi.org/10.3109/14992027.2015.1020971>
- Landsberger, D. M., & Srinivasan, A. G. (2009). Virtual channel discrimination is improved by current focusing in cochlear implant recipients. *Hearing Research*, 254(1–2), 34–41. <https://doi.org/10.1016/j.heares.2009.04.007>
- Litvak, L. M., Spahr, A. J., Saoji, A. A., & Fridman, G. Y. (2007). Relationship between perception of spectral ripple and speech recognition in cochlear implant and vocoder listeners. *The Journal of the Acoustical Society of America*, 122(2), 982–991. <https://doi.org/10.1121/1.2749413>
- Loizou, P. C., Poroy, O., & Dorman, M. (2000). The effect of parametric variations of cochlear implant processors on speech understanding. *The Journal of the Acoustical Society of America*, 108(2), 790–802. <https://doi.org/10.1121/1.429612>
- Mens, L. H., & Berenstein, C. K. (2005). Speech perception with mono- and quadrupolar electrode configurations: A crossover study. *Otology and Neurotology*, 26(5), 957–964. <https://doi.org/10.1097/01.mao.0000185060.74339.9d>
- Narne, V. K., Prabhu, P., Van Dun, B., & Moore, B. C. J. (2018). Ripple glide direction discrimination and its relationship to frequency selectivity estimated using notched noise. *Acta Acustica United With Acustica*, 104(6), 1063–1074. <https://doi.org/10.3813/Aaa.919272>
- Noble, J. H., Gifford, R. H., Hedley-Williams, A. J., Dawant, B. M., & Labadie, R. F. (2014). Clinical evaluation of an image-guided cochlear implant programming strategy. *Audiology and Neurotology*, 19(6), 400–411. <https://doi.org/10.1159/000365273>
- O'Brien, E., & Winn, M. B. (2017). *Aliasing of spectral ripples through CI processors: A challenge to the interpretation of correlation with speech recognition scores* [Poster presentation]. Conference on Implantable Auditory Prostheses (CIAP), Lake Tahoe, United States.
- Riss, D., Arnoldner, C., Baumgartner, W. D., Kaider, A., & Hamzavi, J. S. (2008). A new fine structure speech coding strategy: Speech perception at a reduced number of channels. *Otology & Neurotology*, 29(6), 784–788. <https://doi.org/10.1097/MAO.0b013e31817fe00f>
- Saoji, A. A., Litvak, L., Spahr, A. J., & Eddins, D. A. (2009). Spectral modulation detection and vowel and consonant identifications in cochlear implant listeners. *The Journal of the Acoustical Society of America*, 126(3), 955–958. <https://doi.org/10.1121/1.3179670>
- Schvartz-Leyzac, K. C., Zwolan, T. A., & Pfungst, B. E. (2017). Effects of electrode deactivation on speech recognition in multichannel cochlear implant recipients. *Cochlear Implants International*, 18(6), 324–334. <https://doi.org/10.1080/14670100.2017.1359457>
- Spahr, A. J., Dorman, M. F., Litvak, L. M., Van Wie, S., Gifford, R. H., Loizou, P. C., Losielle, L. M., Oakes, T., & Cook, S. (2012). Development and validation of the AzBio sentence lists. *Ear and Hearing*, 33(1), 112–117. <https://doi.org/10.1097/AUD.0b013e31822c2549>
- Supin, A., Popov, V. V., Milekhina, O. N., & Tarakanov, M. B. (1994). Frequency resolving power measured by rippled noise. *Hearing Research*, 78(1), 31–40. [https://doi.org/10.1016/0378-5955\(94\)90041-8](https://doi.org/10.1016/0378-5955(94)90041-8)
- van den Honert, C., & Kelsall, D. C. (2007). Focused intracochlear electric stimulation with phased array channels. *The Journal of the Acoustical Society of America*, 121(6), 3703–3716. <https://doi.org/10.1121/1.2722047>
- van Hoesel, R., Böhm, M., Pesch, J., Vandali, A., Battmer, R. D., & Lenarz, T. (2008). Binaural speech unmasking and localization in noise with bilateral cochlear implants using envelope and fine-timing based strategies. *The Journal of the Acoustical Society of America*, 123(4), 2249–2263. <https://doi.org/10.1121/1.2875229>
- Won, J. H., Drennan, W. R., Nie, K., Jameyson, E. M., & Rubinstein, J. T. (2011). Acoustic temporal modulation detection and speech perception in cochlear implant listeners. *The Journal of the Acoustical Society of America*, 130(1), 376–388. <https://doi.org/10.1121/1.3592521>
- Won, J. H., Drennan, W. R., & Rubinstein, J. T. (2007). Spectral-ripple resolution correlates with speech reception in noise in cochlear implant users. *Journal of the Association for Research in Otolaryngology*, 8(3), 384–392. <https://doi.org/10.1007/s10162-007-0085-8>