

S13 Table

Positive and Negative Agreement for CSF Specimens

Assay	Target	Positive Agreement ^a (%)	Negative Agreement ^b (%)
ImmuView	<i>S. pneumoniae</i>	96.7 (81.6 to 100)/30	97.7 (95.1 to 98.9)/300
BinaxNOW		92.3 (64.2 to 100)/13	98.9 (95.9 to 99.9)/187

a, mean (95% CI)/total positive specimens (meningitis and contrived meningitis); b, mean (95%

CI)/total negative CSF specimens (contrived).